



EC-Declaration of Conformity

We **MOLLET Füllstandtechnik GmbH**

Hans-Ulrich-Breymann-Str. 10
D-74706 Osterburken
Tel. +49 (0) 6291 64400 Fax +49 (0) 6291 9846

declare that the devices:

Rotary blade level indicator

Type **DF..A3...** and **DF..A4...**

to which this declaration relates is in accordance with the following standards or other normative documents:

Directive 94/9/EC (Directive ATEX)

Directive 89/336/EEC (Directive EMC)

Directive 73/23/EEC (Electrical equipment designed for use within certain voltage limits)

and

EN 50281-1-1:2000, EN 50014:1997+A1+A2,

EN 50018:2000+A1, EN 50019:2000,

EN 1127-1:1997, EN 13463-1:2001, prEN 13463-5:2003,

EN 50281-1-2, CLC/TR 50404:2003

EN 55022, EN 55024, EN 61000-6-2

EN 1050, EN ISO 12100-1, EN ISO 12100-2,

EN 61010-1, EN 60204-1:1997

The Essential Health and Safety Requirements are assured by compliance with the directive and observance of the normative documents.

The marking of the equipment includes the following:

for the electrical part:

CE ₀₀₃₂ **Ex** II 1D2G/2GD EEx de IIB T6 T 70 °C IP 66
-20 °C ≤ T_a ≤ +60 °C

EC-type-examination Certificate Number **IBExU04ATEX1033X**

for the non-electrical part:

CE ₀₀₃₂ **Ex** II 1D2G/2GD c IIB T X

EC-type-examination Certificate Number **IBExU04ATEX1001X**

Osterburken, March 04, 2004

Production quality assurance:

**Qualitäts-
Management**

Regelmäßige freiwillige
Überwachung nach
DIN EN ISO 9001:2000



**Qualitäts-
Sicherungs-
System**

Anerkennung nach
Richtlinie 94/9/EG



Wolfgang Hageleit